

URGENT: FIELD SAFETY NOTICE	
Description	Field Safety Corrective Action for specific lots of DAILIES TOTAL1® and TOTAL30® contact lenses
Product Reference	DAILIES TOTAL1®, TOTAL30®

August XX, 2024

Dear Valued Customer,

Alcon has initiated a Field Safety Corrective Action for specific lots of DAILIES TOTAL1® and TOTAL30® contact lenses. An isolated quality issue was identified with a material supplied by a third-party vendor that was used in the production of contact lenses at a single manufacturing site. As a result, the affected lenses may not meet Alcon quality and/or performance standards for the entirety of their labeled shelf life.

The contact lenses from the identified lots are not expected to cause an increased risk to the wearer. As with any contact lens, there is a remote likelihood that wearing a lens from the affected lots may lead to temporary issues such as irritation, redness, or blurred vision. These symptoms typically resolve after lens removal and/or replacement (as directed in the patient leaflet).

According to our records, affected contact lens lots have been shipped to you. Please discontinue use of the product and contact the retailer where you purchased the contact lenses for further information on next steps.

Sincerely,